



LAB N° 0032 L

XXXXXXXXXXXXXXXXXXXX  
XXXXXXXXXXXXXXXXXXXX  
XXXXXXXXXXXXXXXXXXXX  
IT

Client Account Number: A00963855DQX  
Eurofins Quote Number: PO9YPH21001801

Eurofins Sample Number LV21AA1034-1		
<b>Original Received Date:</b>	20-Jan-2021	
<b>Description:</b>	§ BMP30-NWF-SB + PhotoACTIVE® TEX	
<b>Lot Number:</b>	§ 1	
Analysis	Result	Unit
<b># Citotossicità in vitro/In vitro cytotoxicity</b>		
RESULT:	Not cytotoxic	----
Test item:	30mm <sup>2</sup>	----
Vehicle:	Routine medium	----
Positive control:	30mm <sup>2</sup> of Latex	----
Negative control:	30mm <sup>2</sup> of HDPE	----
Notes:	N/A	----
<b>Addendum #1: Qualitative and quantitative evaluation</b>		
Method: ISO 10993-5:2009		
Analysis Date: 01-Feb-2021 to 03-Feb-2021		



LAB N° 0032 L

**Supplemental Information**

Analytical Report Version 2 Comment: The present test report LV21AA1034-1 Vers 2 supersedes the test report LV21AA1034-1 Vers 1 in order to modify the name of the sample, as requested by the Sponsor.

Sample description: white cloth


Storage: room temperature

# # The Eurofins Testing Laboratory is accredited by ACCREDIA only for tests marked with #. This test report may not be reproduced in part unless expressly approved in writing by Eurofins Biolab S.r.l. The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor. Information provided by the Sponsor are under Sponsor responsibility, and are marked by the symbol "§".

**Contracted Company: Eurofins Biolab Srl (Vimodrone)**

Via Bruno Buozzi, 2, Vimodrone, MI 20055 Italy  
InfoFarma@eurofins.com

*This Certificate of Analysis cannot be modified and/or partially reproduced without the written authorization of the laboratory.  
Questions about this report should be directed to your project manager or the general email listed above.*

	In vitro cytotoxicity ISO10993-5:2009	1-P-QM-TEM-9070229
	Test by Direct Contact - NRU	Addendum N. 1

Study ID: // Sample ID: LV21AA1034-1

Test start: 01/02/2021 Test end: 03/02/2021

Cell line	Manufacturer	
Mammal fibroblasts BALB/3T3 clone A31 (ATCC® CCL163™)	American Type Culture Collection (ATCC)	
Reagent	Manufacturer	Batch
Dulbecco's Modified Eagle Medium (DMEM)	Lonza	0000920735
Fetal Bovine Serum (FBS)	Sigma-Aldrich	20C567
Penicillin/Streptomycin solution	Sigma-Aldrich	0000088375

**QUALITATIVE EVALUATION**

		Contact time: 24 h						Contact time: 24 h			
		PLATE 1						PLATE 2			
		Blank	Vehicle control	Negative control	Test Sample			Blank	Vehicle control	Positive control	Test Sample
Replicate 1			0	0	0	Replicate 1			0	4	0
Replicate 2			0	0	0	Replicate 2			0	4	0
Replicate 3			0	0	0	Replicate 3			0	4	0

**INTERPRETATION OF RESULT**

Grade	Reactivity	Conditions of all Cultures
0	None	No detectable zone around or under specimen
1	Slight	Some malformed or degenerated cells under specimen
2	Mild	Zone limited to area under specimen
3	Moderate	Zone extending specimen size up to 1.0 cm
4	Severe	Zone extending farther than 1.0 cm beyond specimen


**ACCEPTANCE CRITERIA**

Negative control	
Grade ≤1	VALID

Positive control	
Grade ≥3	VALID

**RESULTS**

	Reactivity grade
Test Sample	0

	In vitro cytotoxicity ISO10993-5:2009	1-P-QM-TEM-9070229
	Test by Direct Contact - NRU	Addendum N. 1

Study ID: // Sample ID: LV21AA1034-1

Test start: 01/02/2021 Test end: 03/02/2021

**QUANTITATIVE EVALUATION - 540 nm**

	Contact time: 24 h				Contact time: 24 h				
	PLATE 1				PLATE 2				
	Blank	Vehicle control	Negative control	Test Sample	Blank	Vehicle control	Positive control	Test Sample	
Replicate 1	0,051	1,131	1,117	1,100	Replicate 1	0,049	1,100	0,046	1,107
Replicate 2	0,052	1,116	1,096	1,103	Replicate 2	0,051	1,110	0,049	1,170
Replicate 3	0,045	1,154	1,164	1,195	Replicate 3	0,050	1,153	0,044	1,174

**MEAN, STANDARD DEVIATION, CV% AND VIABILITY**

	Mean OD	Standard Deviation	CV %	Mean OD - Mean OD Blanks	Viability %
Blanks	0,050				
Vehicle control	1,127	0,023	2,007	1,078	100
Negative control	1,126	0,035	3,093	1,076	100
Positive control	0,046	0,003	5,432	-0,003	0
Test Sample	1,142	0,043	3,743	1,092	101

**ACCEPTANCE CRITERIA**

	Vehicle ≥ 0,3			
Mean OD	VALID			
	Negative control ≥ 70%	Positive Control < 70%		
Quantitative Evaluation	VALID	VALID		
	Vehicle control	Negative control	Positive control	Test Sample
CV between replicates ≤ 15%	VALID	VALID	VALID	VALID

$$% \text{ viability} = \frac{OD_{\text{mean sample}} - OD_{\text{mean blank}}}{OD_{\text{mean vehicle}} - OD_{\text{mean blank}}} \times 100$$

**INTERPRETATION OF RESULTS**

Reduction of Viability	Result
≤ 30%	Not Cytotoxic
> 30%	Cytotoxic

**RESULTS**

	Reduction Viability %	
Test Sample	0	NOT CYTOTOXIC


\*0 (% viability test item ≥ vehicle); 100 (% viability test item ≤ Blank)

Finished on: 03/02/21

Technician signature: *[Signature]*

Date: 04/02/21

Approved by: *[Signature]*

	<b>Medical Device Testing</b>	<b>Test Facility</b> Eurofins Biolab S.r.l.	<b>Sample ID:</b> LV21AA1034-1
			<b>Page:</b> 1 of 1

ADDENDUM N.2: EXPERIMENTAL REPORT			
<b>REFERENCE/GUIDELINE:</b>	ISO 10993-5:2009 - In vitro cytotoxicity		
<b>CELL LINE</b>	Mammal fibroblasts BALB/3T3 clone A31 (ATCC®; CCL163™) Source: ATCC.		
<b>MATERIALS</b>	Fetal Bovine Serum (FBS), Neutral Red dye, Trypan Blue, Penicillin/Streptomycin solution, Dulbecco's Phosphate buffer solution (DPBS)	Sigma-Aldrich	
	Dulbecco's Modification of Eagle's Medium (DMEM), Trypsin-EDTA	Lonza	
	Acetic Acid, Ethanol solution	VWR	
	Water for Injection	Eurospital	
	High density polyethylene (HDPE, USP Reference Standard negative control)	Nova Chimica (lot.K0M357)	
	Latex from laboratory gloves	Artsana (lot.1811191C170)	
<b>EQUIPMENT</b>	Laminar flow hood, CO <sub>2</sub> incubator, Microplate reader Mod EL800, Chronometer, Common laboratory equipment, Water, Inverted Microscope Diavert, Orbital shaker, Refrigerator		
<b>EXPERIMENTAL DESIGN</b>			
The experimental design included two 12-well plate containing a subconfluent cell monolayer subdivided in the following groups:			
<b>GROUPS</b>	<b>REPLICATES PLATE N.1</b>		
	Blank	Blank	Blank
	Vehicle	Vehicle	Vehicle
	Negative control	Negative control	Negative control
	Test sample	Test sample	Test sample
<b>GROUPS</b>	<b>REPLICATES PLATE N.2</b>		
	Blank	Blank	Blank
	Vehicle	Vehicle	Vehicle
	Positive control	Positive control	Positive control
	Test sample	Test sample	Test sample
<b>BLANK</b>	Supplemented culture medium alone (without cells).		
<b>VEHICLE</b>	Supplemented culture medium (without test sample).		
<b>TEST SAMPLE</b>	30 mm <sup>2</sup> of the test sample were placed in the middle of each well.		
<b>NEGATIVE CONTROL</b>	The negative control was represented by 30 mm <sup>2</sup> of HDPE placed in the middle of each well.		
<b>POSITIVE CONTROL</b>	The positive control was represented 30 mm <sup>2</sup> of latex placed in the middle of each well.		
<b>TREATMENT:</b> Verified that a subconfluent monolayer was present, supplemented culture medium was replaced with 1mL of fresh supplemented culture medium and the test sample was added. The plates were incubated in a thermostat at (37 ± 1)°C in (5±1)% CO <sub>2</sub> atmosphere for 24 hours. This procedure was repeated for vehicle, negative and positive controls.			
<b>QUALITATIVE EVALUATION (GRADE OF CYTOTOXICITY):</b> After 24 hours the plates were observed under an inverted microscope and biological reactions were evaluated following a 0 to 4 scale according to ISO10993-5:2009.			
<b>QUANTITATIVE EVALUATION (OPTICAL DENSITY):</b> After microscopic observation, cells were treated with Neutral Red Medium for 3 hours at (37 ± 1)°C in (5±1)% CO <sub>2</sub> atmosphere. Subsequently, the Neutral Red medium was removed and each well was rinsed with DPBS. The plates were totally made dry reversing the plates, then Desorb Solution was added and the plates were incubated for at least 15 minutes at room temperature with gentle agitation to form a homogeneous solution. Optical density was measured at 540nm by Gen5 software (Biotek) using microtiter plate reader.			
$\% \text{ of cell viability} = \frac{\text{OD test sample} - \text{OD blank}}{\text{OD vehicle} - \text{OD blank}} \cdot 100$			
<b>ACCEPTABILITY CRITERIA</b>	QUALITATIVE EVALUATION	Negative control ≤ 1; Positive control ≥ 3	
	QUANTITATIVE EVALUATION	The OD mean of the vehicle must be ≥ 0,3. The positive control % cellular viability must be < 70%. The negative control % cellular viability must be ≥ 70%. Coefficient of variation of each group must be ≤15%.	
<b>INTERPRETATION OF RESULTS</b>	The achievement of a numerical grade greater than 2 is considered a cytotoxic effect. A cellular viability reduction more than 30% is considered a cytotoxic effect.		
<b>QUALITY CRITERIA</b>	Satisfied		
<b>QUANTITATIVE EVALUATION</b>	Cells treated with test sample have shown a cell viability reduction of <b>0±0,76%</b> .		

This test report may not be reproduced in part unless expressly approved in writing by Eurofins Biolab S.r.l. The test results relate only to the tested items. Sampling, except specific indication on test report, is always intended to be made by the Sponsor. Information provided by the Sponsor are under Sponsor responsibility.

### Eurofins Biolab S.r.l.

Società con Socio unico sottoposta  
a direzione e coordinamento della società  
Eurofins Pharma Services Italia Holding  
parte di Eurofins Scientific Group  
InfoFarma@eurofins.com  
www.eurofins.it www.biolab.it

Via Bruno Buozzi, 2  
20055 Vimodrone (MI)  
Tel. + 39-022507151  
Fax + 39-0225071599  
certificata@pec.biolab.it

C.SOC. € 100.000 i.v.  
P. IVA 00762140960  
C.F. 03765750157  
REA MI 966696  
D-U-N-S 429117112  
CIT005

Reviewed and electronically signed for Technical Supervisor Approval by  
Elena Bizzotto, Employee  
for Eurofins Biolab Srl, on 18-Feb-2021 15:47:57 UTC+01:00